



ISO/CASCO CLARIFICATION REQUEST FORM

Date of submission: 12 April 2018

1. Requesting ISO Member or A liaison member: IFIA
2. Contact person: Samuel Hill
3. Position: Junior Accreditation Manager
4. Email address: shill@ifia-federation.org
5. Please specify the ISO/CASCO document by name and number (ISO/IEC 17XXX) and clause number: ISO/IEC 17025:2017 8.5.2
6. Clarification request, please formulate the request clearly and where possible in a format that enables a YES or NO answer: ISO/IEC 17025 Section 1 (Scope) states that ISO/IEC 17025 “specifies general requirements for the competence, impartiality and consistent operation of laboratories.” The standard itself thus establishes that all requirements therein pertain to the competence, impartiality and consistent operation of laboratories. ISO/IEC 17025 clause 8.5.2 requires laboratories to plan actions to address risks and opportunities. ISO/IEC 17025 clause 8.9.3 requires the output of the management review to record all actions related to improvement of the laboratory activities. The actions planned as a result of requirements in 8.5.2 are included in the record required in 8.9.3 and both relate to the competence, impartiality and consistent operation of the laboratory. Is this correct?
7. Consensus position of the maintenance group (This section is only to be completed by the maintenance group members) The clarifications requested appear to be hypothetical as insufficient information has been provided related to these specific problems associated with the implementation of the standard. Furthermore, the Working Group’s intention of limiting the impact of “Risk” only relates to the methodology applied. “Risk Management” and the implementation of a formal risk management system (e.g, ISO 31000) is not a requirement.

Send the filled form to casco@iso.org